
Midmark M11 - v1.0.0
Total Cycles: 5475

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2025 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.9 F 0.0
2:00 177.5 F 3.5
4:00 186.8 F 7.5
6:00 204.2 F 11.5
8:00 221.4 F 4.9
10:00 225.9 F 9.3
12:00 239.2 F 11.5
14:00 239.2 F 10.6
16:00 235.8 F 8.2
18:00 248.5 F 14.6
20:00 259.2 F 20.5
22:00 269.2 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.8 F 29.6
2:00 272.6 F 29.2
3:00 271.4 F 28.8
4:00 272.1 F 29.5

Min 271.3 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 228.2 F 0.0
10:00 232.8 F 0.0
15:00 227.6 F 0.0
20:00 230.4 F 0.0
25:00 230.7 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 23:56
STERILIZING: 4:00
VENTING: 3:60
DRYING: 30:00
TOTAL CYCLE: 01:03:18
CYCLE COMPLETE

04/03/2025 01:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____