
Midmark M11 - v1.0.0
Total Cycles: 1292

Name: RITTER 1
Sterilizer ID: V2178198
10/22/2021 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.5 F 0.0
2:00 195.2 F 3.2
4:00 207.6 F 7.8
6:00 218.0 F 13.5
8:00 234.4 F 8.9
10:00 241.0 F 10.7
12:00 243.3 F 12.1
14:00 241.1 F 10.7
16:00 254.2 F 17.3
18:00 265.7 F 24.2
20:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.8 F 28.5
2:00 272.1 F 28.8
3:00 272.7 F 29.0
4:00 271.9 F 29.0

Min 271.2 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 224.7 F 0.0
10:00 211.7 F 0.0
15:00 203.3 F 0.0
20:00 204.6 F 0.0
25:00 201.6 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 20:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

10/22/2021 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____