
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 000300
13:08 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 2:03
STERILIZING 3:10
136.0 C 228kPa 3:10
136.4 C 231kPa 3:40
136.4 C 233kPa 4:10
136.4 C 232kPa 4:40
136.2 C 232kPa 5:10
136.3 C 232kPa 5:40
136.3 C 232kPa 6:10
136.3 C 233kPa 6:40
136.3 C 233kPa 6:40
VENTING 6:40
AIR DRYING 7:36
CYCLE COMPLETE 22:36

Digital Signature #
37EBC98C39034869
