
Midmark M11 - v1.0.5
Total Cycles: 487

Name: RITTER2
Sterilizer ID: V2313228
10/27/2021 06:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 80.0 F 0.0
4:00 85.4 F 0.0
6:00 97.7 F 0.1
8:00 142.2 F 0.2
10:00 209.6 F 0.4
12:00 222.7 F 3.8
14:00 228.1 F 7.1
16:00 237.7 F 10.8
18:00 232.7 F 8.5
20:00 243.4 F 12.2
22:00 239.5 F 10.2
24:00 235.8 F 8.7
26:00 250.7 F 15.7
28:00 262.0 F 22.2
30:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.7 F 29.6
2:00 271.7 F 29.6
3:00 272.7 F 29.4
4:00 272.0 F 29.7

Min 271.4 F 29.1
Max 273.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 234.7 F 0.0
10:00 245.0 F 0.0
15:00 232.1 F 0.0
20:00 224.1 F 0.0
25:00 225.2 F 0.0
30:00 199.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 31:16
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:09:16
CYCLE COMPLETE

10/27/2021 07:42 AM
ACCEPTED: X
REJECTED:
Operator: U01