

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 003703  
9:28 27/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:58  
PRESSURIZING 2:04  
STERILIZING 3:09  
136.4 C 222kPa 3:09  
136.3 C 224kPa 3:39  
136.4 C 224kPa 4:09  
136.0 C 222kPa 4:39  
136.2 C 220kPa 5:09  
136.4 C 223kPa 5:39  
136.3 C 222kPa 6:09  
136.4 C 223kPa 6:39  
136.4 C 223kPa 6:39  
VENTING 6:39  
AIR DRYING 7:28  
CYCLE COMPLETE 7:30

-----  
Digital Signature #  
32A114FA446FDF07  
-----