
Midmark M11 - v1.0.0
Total Cycles: 931

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2021 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.8 F 0.0
2:00 176.9 F 2.1
4:00 183.8 F 5.5
6:00 201.6 F 9.7
8:00 221.6 F 7.8
10:00 226.0 F 9.0
12:00 236.3 F 10.8
14:00 239.2 F 11.4
16:00 237.3 F 9.3
18:00 250.6 F 15.5
20:00 261.5 F 21.7
22:00 271.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 271.8 F 28.5
2:00 272.5 F 28.8
3:00 271.7 F 29.1
4:00 271.2 F 28.6

Min 271.2 F 28.4
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 215.5 F 0.0
10:00 214.7 F 0.0
15:00 213.1 F 0.0
20:00 215.0 F 0.0
25:00 220.2 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

06/23/2021 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____