
Midmark M11 - v1.0.0
Total Cycles: 1661

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2022 01:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.8 F 0.0
2:00 177.3 F 2.6
4:00 195.1 F 7.0
6:00 212.5 F 12.8
8:00 224.6 F 7.2
10:00 237.7 F 12.6
12:00 237.7 F 10.0
14:00 236.5 F 8.4
16:00 250.1 F 15.3
18:00 262.2 F 22.2
20:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.8
2:00 271.6 F 28.3
3:00 272.7 F 28.9
4:00 271.6 F 28.2

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 219.1 F 0.0
10:00 233.4 F 0.0
15:00 225.9 F 0.0
20:00 222.4 F 0.0
25:00 224.1 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

02/23/2022 02:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____