
Midmark M11 - v1.0.5
Total Cycles: 3119

Name: RITTER2
Sterilizer ID: V2313228
11/13/2023 04:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.8 F 0.0
2:00 156.7 F 2.4
4:00 173.9 F 6.0
6:00 195.3 F 10.6
8:00 214.6 F 5.1
10:00 223.9 F 9.7
12:00 232.8 F 9.0
14:00 240.5 F 12.9
16:00 242.0 F 12.0
18:00 254.7 F 18.5
20:00 265.5 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.0 F 29.9
2:00 272.8 F 30.1
3:00 272.5 F 29.8
4:00 272.0 F 29.6

Min 271.1 F 29.0
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.5 F 0.0
5:00 194.6 F 0.0
10:00 195.9 F 0.0
15:00 195.7 F 0.0
20:00 194.9 F 0.0
25:00 193.5 F 0.0
30:00 191.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:49
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

11/13/2023 05:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____