
Midmark M11 - v1.0.0
Total Cycles: 3158

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2023 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 182.8 F 0.0
4:00 200.2 F 0.3
6:00 220.1 F 5.4
8:00 232.9 F 10.3
10:00 235.3 F 7.7
12:00 243.8 F 13.1
14:00 243.0 F 11.8
16:00 241.4 F 10.8
18:00 255.1 F 17.8
20:00 266.9 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.4 F 28.5
2:00 271.9 F 28.4
3:00 272.3 F 28.4
4:00 272.0 F 28.4

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 218.0 F 0.0
10:00 232.8 F 0.0
15:00 226.9 F 0.0
20:00 226.6 F 0.0
25:00 229.1 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

04/25/2023 09:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____