
Midmark M11 - v1.0.0
Total Cycles: 2408

Name: RITTER 1
Sterilizer ID: V2178198
10/06/2022 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.6 F 0.0
2:00 107.0 F 0.0
4:00 146.9 F 1.0
6:00 168.8 F 3.2
8:00 185.0 F 6.5
10:00 201.8 F 10.5
12:00 216.1 F 5.4
14:00 225.3 F 7.8
16:00 234.6 F 11.9
18:00 233.4 F 8.1
20:00 243.9 F 13.5
22:00 241.5 F 11.4
24:00 253.8 F 17.5
26:00 264.5 F 23.9
28:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.5
2:00 271.8 F 28.7
3:00 272.0 F 29.2
4:00 272.4 F 29.2

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 193.9 F 0.0
10:00 198.6 F 0.0
15:00 211.9 F 0.0
20:00 239.0 F 0.0
25:00 234.6 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:06
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:06:13
CYCLE COMPLETE

10/06/2022 10:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____