

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000740  
9:53 23/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:09  
PRESSURIZING 2:18  
STERILIZING 3:16  
136.0 C 229kPa 3:16  
136.2 C 232kPa 3:46  
136.3 C 233kPa 4:16  
136.4 C 233kPa 4:46  
136.3 C 232kPa 5:16  
136.4 C 233kPa 5:46  
136.4 C 233kPa 6:16  
136.4 C 233kPa 6:46  
136.4 C 233kPa 6:46  
VENTING 6:47  
AIR DRYING 7:38  
CYCLE COMPLETE 7:47  
-----

Digital Signature #  
A3ACC1E5428AC319  
-----