
Midmark M11 - v1.0.0
Total Cycles: 830

Name: RITTER 1
Sterilizer ID: V2178198
05/20/2021 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.9 F 0.0
2:00 172.5 F 2.3
4:00 186.5 F 7.3
6:00 204.2 F 13.4
8:00 222.9 F 9.6
10:00 231.6 F 7.6
12:00 240.1 F 13.2
14:00 242.5 F 11.5
16:00 254.9 F 18.6
18:00 267.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 273.1 F 29.6
2:00 271.8 F 28.8
3:00 273.5 F 29.8
4:00 272.6 F 29.2

Min 271.5 F 28.8
Max 273.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 227.2 F 0.0
10:00 237.3 F 0.0
15:00 228.6 F 0.0
20:00 225.7 F 0.0
25:00 227.6 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:25
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:57:60
CYCLE COMPLETE

05/20/2021 12:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____