
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001541
15:35 10/28/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 5:31
STERILIZING 7:00
136.0 C 220kPa 7:00
136.4 C 220kPa 7:30
136.3 C 222kPa 8:00
136.4 C 223kPa 8:30
136.1 C 220kPa 9:00
136.5 C 221kPa 9:30
136.4 C 223kPa 10:00
136.5 C 220kPa 10:30
VENTING 10:30
AIR DRYING 11:34
CYCLE COMPLETE 21:46

Digital Signature #
0DEE1BB63C7F52FE
