
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002659
13:47 07/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:25
STERILIZING 3:32
136.4 C 226kPa 3:32
136.4 C 222kPa 4:02
136.4 C 226kPa 4:32
136.3 C 224kPa 5:02
136.3 C 225kPa 5:32
136.4 C 225kPa 6:02
136.3 C 225kPa 6:32
136.4 C 226kPa 7:02
136.4 C 226kPa 7:02
VENTING 7:02
AIR DRYING 7:53
CASSETTE REMOVED
CYCLE COMPLETE 16:24

Digital Signature #
854426FE4FEF3B78
