
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000115
13:51 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 4:31
STERILIZING 5:53
136.4 C 225kPa 5:53
136.3 C 223kPa 6:23
136.3 C 224kPa 6:53
136.4 C 224kPa 7:23
136.4 C 226kPa 7:53
136.4 C 226kPa 8:23
136.4 C 224kPa 8:53
136.4 C 226kPa 9:23
136.4 C 226kPa 9:23
VENTING 9:23
AIR DRYING 10:29
CYCLE COMPLETE 25:30

Digital Signature #
45A70F4D9B4F1E64
