
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005525
9:30 15/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:31
STERILIZING 3:15
136.0 C 220kPa 3:15
136.4 C 224kPa 3:45
136.2 C 222kPa 4:15
136.2 C 223kPa 4:45
136.4 C 225kPa 5:15
136.4 C 224kPa 5:45
136.3 C 224kPa 6:15
136.4 C 222kPa 6:45
136.4 C 222kPa 6:45
VENTING 6:45
AIR DRYING 7:38
CYCLE COMPLETE 11:50

Digital Signature #
430EBDCD3D253025
