
Midmark M11 - v1.0.0
Total Cycles: 5027

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2024 03:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.2 F 0.0
2:00 134.2 F 0.3
4:00 154.5 F 0.7
6:00 196.6 F 3.6
8:00 222.8 F 7.8
10:00 222.0 F 12.2
12:00 224.3 F 6.5
14:00 235.8 F 11.3
16:00 234.7 F 8.2
18:00 243.6 F 11.5
20:00 245.7 F 12.8
22:00 257.4 F 19.2
24:00 268.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.8
2:00 271.4 F 28.1
3:00 272.4 F 28.9
4:00 271.9 F 28.4

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.2 F 0.0
5:00 224.1 F 0.0
10:00 192.9 F 0.0
15:00 191.4 F 0.0
20:00 240.8 F 0.0
25:00 235.6 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:19
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:53
CYCLE COMPLETE

10/25/2024 04:15 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____