

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3588

Name: RITTER2  
Sterilizer ID: V2313228  
05/03/2024 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 116.6 F 0.0  
2:00 116.6 F 1.0  
4:00 130.1 F 3.1  
6:00 159.6 F 6.1  
8:00 183.0 F 9.5  
10:00 197.0 F 11.2  
12:00 221.2 F 7.2  
14:00 225.5 F 11.4  
16:00 230.7 F 7.3  
18:00 238.8 F 12.6  
20:00 239.8 F 10.6  
22:00 251.4 F 16.5  
24:00 261.8 F 22.6  
26:00 272.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.7  
1:00 273.6 F 30.2  
2:00 271.4 F 29.7  
3:00 272.6 F 30.2  
4:00 271.4 F 29.6

Min 271.3 F 29.6  
Max 273.6 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.4 F 0.0  
5:00 209.9 F 0.0  
10:00 195.1 F 0.0  
15:00 193.0 F 0.0  
20:00 189.5 F 0.0  
25:00 186.8 F 0.0  
30:00 185.5 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 27:58  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:06:34  
CYCLE COMPLETE

05/03/2024 01:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_