

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1868

Name: RITTER 1  
Sterilizer ID: V2178198  
04/26/2022 07:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.7 F 0.0  
2:00 87.7 F 0.0  
4:00 91.2 F 0.0  
6:00 118.6 F 0.2  
8:00 151.7 F 1.6  
10:00 175.1 F 3.7  
12:00 205.1 F 6.6  
14:00 215.3 F 9.9  
16:00 226.1 F 13.5  
18:00 225.3 F 7.0  
20:00 235.6 F 11.1  
22:00 232.4 F 7.2  
24:00 245.7 F 13.2  
26:00 243.3 F 11.7  
28:00 255.9 F 18.3  
30:00 266.9 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.2  
1:00 272.5 F 28.9  
2:00 271.5 F 28.1  
3:00 272.4 F 28.7  
4:00 271.5 F 28.3  
  
Min 271.4 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.9 F 0.0  
5:00 236.0 F 0.0  
10:00 222.0 F 0.0  
15:00 236.4 F 0.0  
20:00 223.5 F 0.0  
25:00 226.3 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 31:21  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:09:29  
CYCLE COMPLETE

04/26/2022 08:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_