
Midmark M11 - v1.0.5
Total Cycles: 591

Name: RITTER2
Sterilizer ID: V2313228
11/30/2021 04:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.2 F 0.0
2:00 145.9 F 2.6
4:00 167.8 F 6.2
6:00 193.8 F 11.0
8:00 213.1 F 5.1
10:00 226.7 F 10.7
12:00 229.4 F 7.5
14:00 241.7 F 12.5
16:00 243.1 F 12.7
18:00 255.7 F 19.2
20:00 266.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 271.9 F 29.3
2:00 272.1 F 29.4
3:00 272.3 F 29.7
4:00 272.7 F 30.1

Min 271.6 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.8 F 0.0
5:00 230.0 F 0.0
10:00 227.3 F 0.0
15:00 229.5 F 0.0
20:00 228.4 F 0.0
25:00 226.5 F 0.0
30:00 233.2 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

12/01/2021 07:26 AM
ACCEPTED: X
REJECTED:
Operator: U02