
Midmark M11 - v1.0.5
Total Cycles: 862

Name: RITTER2
Sterilizer ID: V2313228
03/04/2022 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 78.8 F 0.0
4:00 93.9 F 0.0
6:00 134.8 F 0.3
8:00 164.3 F 2.6
10:00 190.7 F 6.7
12:00 210.5 F 11.5
14:00 220.3 F 5.5
16:00 232.7 F 10.3
18:00 233.2 F 7.9
20:00 244.3 F 13.2
22:00 243.0 F 12.0
24:00 256.5 F 19.1
26:00 268.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.2 F 29.4
2:00 272.3 F 29.1
3:00 271.8 F 28.6
4:00 271.8 F 29.1

Min 271.6 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.0 F 0.0
5:00 205.9 F 0.0
10:00 211.5 F 0.0
15:00 216.6 F 0.0
20:00 220.0 F 0.0
25:00 223.9 F 0.0
30:00 219.9 F 0.0

Drying Complete

FILLING: 1:26
HEATING: 27:04
STERILIZING: 4:00
VENTING: 2:45
DRYING: 30:00
TOTAL CYCLE: 01:05:16
CYCLE COMPLETE

03/04/2022 02:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____