
Midmark M11 - v1.0.0
Total Cycles: 5008

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2024 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 73.5 F 0.0
4:00 107.4 F 0.0
6:00 121.1 F 0.1
8:00 144.7 F 0.3
10:00 174.3 F 0.9
12:00 197.3 F 3.8
14:00 208.6 F 6.9
16:00 220.7 F 10.4
18:00 226.4 F 8.5
20:00 231.7 F 8.6
22:00 242.3 F 13.2
24:00 239.8 F 10.4
26:00 236.0 F 7.9
28:00 249.7 F 14.7
30:00 261.0 F 21.0
32:00 271.3 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 27.8
1:00 272.3 F 28.4
2:00 272.1 F 28.0
3:00 271.9 F 28.2
4:00 272.6 F 28.4

Min 271.5 F 27.7
Max 272.8 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 197.8 F 0.0
10:00 203.5 F 0.0
15:00 210.5 F 0.0
20:00 208.3 F 0.0
25:00 206.9 F 0.0
30:00 207.5 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:32
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:10:35
CYCLE COMPLETE

10/18/2024 09:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____