
Midmark M11 - v1.0.0
Total Cycles: 1792

Name: RITTER 1
Sterilizer ID: V2178198
04/05/2022 07:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 75.7 F 0.0
4:00 87.8 F 0.0
6:00 114.0 F 0.1
8:00 144.3 F 1.0
10:00 173.4 F 4.3
12:00 195.5 F 7.9
14:00 217.2 F 11.4
16:00 223.2 F 4.5
18:00 228.6 F 8.4
20:00 238.8 F 12.5
22:00 236.5 F 9.1
24:00 241.0 F 10.2
26:00 246.2 F 13.1
28:00 257.6 F 19.3
30:00 268.1 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.5
2:00 272.2 F 28.8
3:00 271.7 F 28.1
4:00 272.6 F 28.9

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 232.1 F 0.0
10:00 189.6 F 0.0
15:00 184.6 F 0.0
20:00 183.2 F 0.0
25:00 182.9 F 0.0
30:00 188.8 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:17
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:09
CYCLE COMPLETE

04/05/2022 08:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____