
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002382
9:02 16/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:14
STERILIZING 4:09
136.0 C 220kPa 4:09
136.4 C 224kPa 4:39
136.4 C 224kPa 5:09
136.4 C 224kPa 5:39
136.3 C 221kPa 6:09
136.4 C 222kPa 6:39
136.4 C 224kPa 7:09
136.4 C 224kPa 7:39
136.4 C 224kPa 7:39
VENTING 7:39
AIR DRYING 8:29
CYCLE COMPLETE 38:29

Digital Signature #
C22A101795C07185
