
Midmark M11 - v1.0.5
Total Cycles: 3845

Name: RITTER2
Sterilizer ID: V2313228
08/12/2024 08:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 78.5 F 0.0
4:00 124.7 F 0.0
6:00 105.9 F 0.3
8:00 151.7 F 3.1
10:00 171.4 F 6.5
12:00 189.5 F 10.1
14:00 205.2 F 9.9
16:00 221.9 F 6.6
18:00 228.2 F 10.2
20:00 233.1 F 9.7
22:00 238.1 F 10.6
24:00 234.9 F 8.3
26:00 246.4 F 14.0
28:00 257.5 F 20.0
30:00 267.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.1 F 29.0
2:00 272.4 F 29.3
3:00 272.2 F 29.6
4:00 271.6 F 29.5

Min 271.0 F 28.9
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 229.3 F 0.0
10:00 222.6 F 0.0
15:00 189.4 F 0.0
20:00 189.3 F 0.0
25:00 198.8 F 0.0
30:00 188.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:20
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

08/12/2024 09:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____