

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1835

Name: RITTER 1  
Sterilizer ID: V2178198  
04/18/2022 01:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
2:00 196.8 F 4.3  
4:00 203.6 F 9.8  
6:00 223.6 F 5.8  
8:00 231.6 F 10.4  
10:00 233.1 F 7.3  
12:00 244.3 F 12.5  
14:00 245.3 F 12.6  
16:00 258.4 F 19.6  
18:00 269.9 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.1  
1:00 272.5 F 28.9  
2:00 271.7 F 28.1  
3:00 272.3 F 28.9  
4:00 272.3 F 28.4  
  
Min 271.1 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 234.7 F 0.0  
10:00 202.5 F 0.0  
15:00 201.5 F 0.0  
20:00 206.2 F 0.0  
25:00 213.5 F 0.0  
30:00 219.5 F 0.0  
  
Drying Complete

FILLING: 1:19  
HEATING: 18:47  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:13  
CYCLE COMPLETE

04/18/2022 02:39 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_