
Midmark M11 - v1.0.5
Total Cycles: 1873

Name: RITTER2
Sterilizer ID: V2313228
12/21/2022 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.7 F 0.0
2:00 184.7 F 4.0
4:00 196.9 F 8.8
6:00 213.7 F 8.2
8:00 227.6 F 9.7
10:00 233.7 F 8.5
12:00 243.3 F 13.1
14:00 243.4 F 12.2
16:00 256.4 F 18.9
18:00 267.4 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 272.4 F 29.7
2:00 272.1 F 29.1
3:00 271.2 F 28.7
4:00 271.8 F 29.5

Min 271.0 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 223.0 F 0.0
10:00 236.6 F 0.0
15:00 229.8 F 0.0
20:00 227.7 F 0.0
25:00 228.6 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 19:25
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

12/21/2022 02:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____