

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001408  
13:38 09/08/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:40  
PRESSURIZING 5:16  
STERILIZING 6:41  
136.2 C 220kPa 6:41  
136.4 C 224kPa 7:11  
136.3 C 223kPa 7:41  
136.4 C 222kPa 8:11  
136.2 C 221kPa 8:41  
136.4 C 223kPa 9:11  
136.5 C 221kPa 9:41  
136.3 C 223kPa 10:11  
VENTING 10:11  
AIR DRYING 11:11  
CYCLE COMPLETE 11:14  
-----

Digital Signature #  
146758F88EDDCD6  
-----