
Midmark M11 - v1.0.0
Total Cycles: 3348

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2023 09:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.2 F 0.0
2:00 154.3 F 0.0
4:00 161.1 F 0.0
6:00 193.7 F 4.1
8:00 209.3 F 8.7
10:00 227.0 F 12.7
12:00 232.4 F 9.0
14:00 241.0 F 10.8
16:00 241.3 F 12.0
18:00 239.8 F 10.5
20:00 253.3 F 17.2
22:00 264.7 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 271.8 F 28.9
2:00 272.7 F 29.2
3:00 272.0 F 28.6
4:00 271.7 F 28.8

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 221.9 F 0.0
10:00 221.7 F 0.0
15:00 221.2 F 0.0
20:00 220.1 F 0.0
25:00 220.4 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:48
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:57
CYCLE COMPLETE

06/07/2023 10:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____