

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000398  
14:27 25/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:02  
PRESSURIZING 3:06  
STERILIZING 4:11  
136.0 C 229kPa 4:11  
136.2 C 232kPa 4:41  
136.0 C 230kPa 5:11  
136.0 C 230kPa 5:41  
136.4 C 231kPa 6:11  
136.4 C 233kPa 6:41  
136.1 C 231kPa 7:11  
136.1 C 231kPa 7:41  
136.1 C 231kPa 7:41  
VENTING 7:41  
AIR DRYING 8:36  
CYCLE COMPLETE 9:02  
-----

Digital Signature #  
879F3F576B6ABB2C  
-----