
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001544
12:19 24/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:03
STERILIZING 4:07
136.4 C 224kPa 4:07
136.4 C 226kPa 4:37
136.4 C 227kPa 5:07
136.4 C 226kPa 5:37
136.1 C 224kPa 6:07
136.4 C 226kPa 6:37
136.4 C 226kPa 7:07
136.4 C 227kPa 7:37
136.4 C 227kPa 7:37
VENTING 7:37
AIR DRYING 8:35
CYCLE COMPLETE 11:45

Digital Signature #
C9E77764E0A0C159
