
Midmark M11 - v1.0.5
Total Cycles: 3479

Name: RITTER2
Sterilizer ID: V2313228
03/27/2024 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 70.6 F 0.0
4:00 76.2 F 0.0
6:00 105.0 F 0.1
8:00 147.6 F 2.0
10:00 163.5 F 5.1
12:00 189.9 F 8.5
14:00 212.4 F 12.0
16:00 221.4 F 5.1
18:00 229.1 F 9.0
20:00 238.9 F 13.3
22:00 236.4 F 9.9
24:00 237.8 F 9.4
26:00 246.3 F 13.7
28:00 257.5 F 19.8
30:00 267.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.9
1:00 272.6 F 29.5
2:00 271.0 F 28.9
3:00 273.0 F 30.1
4:00 271.8 F 29.3

Min 271.0 F 28.7
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 228.4 F 0.0
10:00 216.9 F 0.0
15:00 214.5 F 0.0
20:00 207.2 F 0.0
25:00 203.1 F 0.0
30:00 192.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:25
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:35
CYCLE COMPLETE

03/27/2024 09:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____