
Midmark M11 - v1.0.5
Total Cycles: 4075

Name: RITTER2
Sterilizer ID: V2313228
11/01/2024 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.0 F 0.0
2:00 155.9 F 1.9
4:00 175.8 F 5.3
6:00 196.8 F 9.2
8:00 212.1 F 13.4
10:00 221.5 F 7.4
12:00 232.6 F 11.9
14:00 235.0 F 8.8
16:00 240.6 F 11.0
18:00 245.6 F 13.3
20:00 256.9 F 19.4
22:00 267.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.7 F 30.0
2:00 271.6 F 29.5
3:00 272.5 F 29.6
4:00 272.0 F 30.0

Min 270.8 F 28.8
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 250.6 F 0.0
10:00 230.7 F 0.0
15:00 219.4 F 0.0
20:00 231.9 F 0.0
25:00 233.7 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 23:24
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:02:39
CYCLE COMPLETE

11/01/2024 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____