

-----  
Midmark M11 - v1.0.5  
Total Cycles: 380

Name: RITTER2  
Sterilizer ID: V2313228  
09/21/2021 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 88.5 F 0.0  
2:00 89.5 F 0.4  
4:00 104.2 F 1.6  
6:00 129.5 F 3.7  
8:00 154.0 F 6.8  
10:00 172.2 F 10.0  
12:00 194.1 F 13.5  
14:00 215.1 F 7.2  
16:00 224.8 F 11.3  
18:00 224.8 F 6.7  
20:00 238.5 F 11.8  
22:00 236.5 F 9.4  
24:00 247.7 F 14.9  
26:00 258.0 F 20.4  
28:00 267.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.4  
1:00 272.8 F 30.0  
2:00 271.6 F 29.6  
3:00 272.7 F 29.7  
4:00 272.3 F 30.1

Min 271.4 F 29.3  
Max 273.3 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 231.9 F 0.0  
10:00 212.8 F 0.0  
15:00 205.1 F 0.0  
20:00 202.1 F 0.0  
25:00 201.4 F 0.0  
30:00 203.1 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 29:24  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:07:31  
CYCLE COMPLETE

09/21/2021 10:30 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01