
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003789
9:04 08/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:20
STERILIZING 3:17
136.0 C 220kPa 3:17
136.1 C 222kPa 3:47
136.1 C 221kPa 4:17
136.2 C 222kPa 4:47
136.4 C 220kPa 5:17
136.4 C 223kPa 5:47
136.3 C 223kPa 6:17
136.1 C 223kPa 6:47
136.1 C 223kPa 6:47
VENTING 6:47
AIR DRYING 7:35
CYCLE COMPLETE 11:00

Digital Signature #
D98042A16FF7830A
