
Midmark M11 - v1.0.5
Total Cycles: 631

Name: RITTER2
Sterilizer ID: V2313228
12/14/2021 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 167.8 F 2.6
4:00 172.5 F 6.3
6:00 191.4 F 11.1
8:00 215.9 F 5.0
10:00 229.1 F 10.1
12:00 233.7 F 8.4
14:00 242.3 F 13.0
16:00 242.1 F 11.8
18:00 254.7 F 18.3
20:00 265.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.6 F 29.7
2:00 271.3 F 29.0
3:00 272.2 F 29.8
4:00 272.1 F 29.3

Min 271.1 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 227.5 F 0.0
10:00 223.4 F 0.0
15:00 218.5 F 0.0
20:00 218.7 F 0.0
25:00 219.1 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

12/14/2021 12:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____