
Midmark M11 - v1.0.0
Total Cycles: 1794

Name: RITTER 1
Sterilizer ID: V2178198
04/05/2022 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 187.5 F 3.3
4:00 194.9 F 8.1
6:00 213.4 F 10.3
8:00 226.8 F 8.7
10:00 237.9 F 10.6
12:00 241.2 F 11.9
14:00 240.4 F 10.6
16:00 254.3 F 17.7
18:00 266.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.0
2:00 271.5 F 28.5
3:00 272.6 F 29.1
4:00 271.6 F 28.4

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 238.3 F 0.0
10:00 226.3 F 0.0
15:00 225.7 F 0.0
20:00 225.2 F 0.0
25:00 227.9 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

04/05/2022 10:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____