

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2481

Name: RITTER2  
Sterilizer ID: V2313228  
06/07/2023 01:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.4 F 0.0  
2:00 159.1 F 2.8  
4:00 177.2 F 6.6  
6:00 197.7 F 11.4  
8:00 216.7 F 6.1  
10:00 231.4 F 11.2  
12:00 230.5 F 8.1  
14:00 240.8 F 11.7  
16:00 243.8 F 12.9  
18:00 256.1 F 19.4  
20:00 267.0 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.6 F 29.9  
2:00 271.1 F 29.4  
3:00 272.5 F 30.0  
4:00 271.3 F 29.3

Min 271.1 F 29.2  
Max 272.6 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.6 F 0.0  
5:00 212.5 F 0.0  
10:00 231.3 F 0.0  
15:00 226.2 F 0.0  
20:00 223.0 F 0.0  
25:00 223.6 F 0.0  
30:00 221.0 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 21:30  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:00:09  
CYCLE COMPLETE

06/07/2023 02:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_