
Midmark M11 - v1.0.5
Total Cycles: 4496

Name: RITTER2
Sterilizer ID: V2313228
04/08/2025 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.7 F 0.0
2:00 98.9 F 0.4
4:00 101.3 F 2.1
6:00 124.6 F 4.6
8:00 161.7 F 7.6
10:00 173.3 F 11.2
12:00 195.3 F 4.0
14:00 208.8 F 6.5
16:00 223.9 F 10.2
18:00 231.2 F 9.6
20:00 235.8 F 11.3
22:00 234.6 F 9.2
24:00 247.1 F 15.1
26:00 257.8 F 20.9
28:00 267.3 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 30.1
1:00 272.7 F 30.9
2:00 272.9 F 30.8
3:00 272.1 F 30.3
4:00 271.5 F 30.0

Min 271.5 F 30.0
Max 273.4 F 31.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 210.2 F 0.0
10:00 212.7 F 0.0
15:00 213.2 F 0.0
20:00 211.5 F 0.0
25:00 203.3 F 0.0
30:00 206.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:37
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:08:27
CYCLE COMPLETE

04/08/2025 01:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____