

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2692

Name: RITTER 1  
Sterilizer ID: V2178198  
12/19/2022 02:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
2:00 196.2 F 3.6  
4:00 212.4 F 8.5  
6:00 227.9 F 7.9  
8:00 234.8 F 10.1  
10:00 237.0 F 8.0  
12:00 244.6 F 12.5  
14:00 245.1 F 12.7  
16:00 257.9 F 19.4  
18:00 269.4 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 273.0 F 29.4  
2:00 272.5 F 28.8  
3:00 271.8 F 28.9  
4:00 273.0 F 29.5  
  
Min 271.7 F 28.5  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 220.1 F 0.0  
10:00 241.9 F 0.0  
15:00 239.9 F 0.0  
20:00 236.2 F 0.0  
25:00 227.9 F 0.0  
30:00 226.8 F 0.0  
  
Drying Complete

FILLING: 2:37  
HEATING: 19:50  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:40  
CYCLE COMPLETE

12/19/2022 03:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_