
Midmark M11 - v1.0.5
Total Cycles: 1877

Name: RITTER2
Sterilizer ID: V2313228
12/22/2022 07:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.8 F 0.0
2:00 86.9 F 0.0
4:00 131.0 F 0.0
6:00 108.5 F 0.0
8:00 183.8 F 0.1
10:00 199.0 F 1.3
12:00 212.6 F 5.0
14:00 223.0 F 8.6
16:00 237.6 F 12.5
18:00 230.7 F 7.5
20:00 241.0 F 12.3
22:00 238.2 F 9.8
24:00 238.3 F 9.0
26:00 248.7 F 14.6
28:00 260.0 F 21.0
30:00 270.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.7 F 29.3
2:00 271.4 F 28.4
3:00 272.4 F 29.4
4:00 272.4 F 28.8

Min 271.0 F 28.3
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.3 F 0.0
5:00 221.7 F 0.0
10:00 200.0 F 0.0
15:00 201.7 F 0.0
20:00 186.5 F 0.0
25:00 189.0 F 0.0
30:00 171.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 30:51
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:08:53
CYCLE COMPLETE

12/22/2022 08:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____