
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000714
7:17 21/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:58
STERILIZING 3:06
136.2 C 228kPa 3:06
136.2 C 230kPa 3:36
136.4 C 232kPa 4:06
136.4 C 233kPa 4:36
136.3 C 231kPa 5:06
136.4 C 231kPa 5:36
136.3 C 230kPa 6:06
136.5 C 233kPa 6:36
136.5 C 233kPa 6:36
VENTING 6:37
AIR DRYING 7:32
CYCLE COMPLETE 7:34

Digital Signature #
6EE0C8EF7BE55094
