

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001403  
12:06 13/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:20  
PRESSURIZING 3:23  
STERILIZING 4:38  
136.0 C 220kPa 4:38  
136.4 C 224kPa 5:08  
136.4 C 225kPa 5:38  
136.2 C 222kPa 6:08  
136.0 C 221kPa 6:38  
136.2 C 222kPa 7:08  
136.4 C 225kPa 7:38  
136.4 C 223kPa 8:08  
136.4 C 223kPa 8:08  
VENTING 8:09  
AIR DRYING 9:02  
CYCLE COMPLETE 24:02  
-----

Digital Signature #  
73D82B943613D82E  
-----