
Midmark M11 - v1.0.5
Total Cycles: 2949

Name: RITTER2
Sterilizer ID: V2313228
09/29/2023 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.9 F 0.0
2:00 167.4 F 0.2
4:00 172.8 F 1.3
6:00 190.9 F 5.1
8:00 207.7 F 9.4
10:00 220.4 F 9.6
12:00 228.8 F 9.5
14:00 235.7 F 10.0
16:00 242.4 F 12.4
18:00 241.5 F 11.3
20:00 253.6 F 17.5
22:00 264.1 F 23.8
24:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.2 F 29.6
2:00 272.3 F 29.2
3:00 271.1 F 28.7
4:00 272.1 F 29.7

Min 270.9 F 28.6
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 227.5 F 0.0
10:00 196.3 F 0.0
15:00 190.4 F 0.0
20:00 185.1 F 0.0
25:00 196.6 F 0.0
30:00 204.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:02:51
CYCLE COMPLETE

09/29/2023 11:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____