
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 000303
15:22 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:47
PRESSURIZING 1:54
STERILIZING 2:50
136.4 C 233kPa 2:50
136.2 C 232kPa 3:20
136.2 C 231kPa 3:50
136.3 C 233kPa 4:20
136.4 C 232kPa 4:50
136.4 C 233kPa 5:20
136.4 C 232kPa 5:50
136.4 C 231kPa 6:20
136.4 C 231kPa 6:20
VENTING 6:20
AIR DRYING 7:08
CYCLE COMPLETE 16:58

Digital Signature #
F1EDCA0F2857501C
