
Midmark M11 - v1.0.5
Total Cycles: 3764

Name: RITTER2
Sterilizer ID: V2313228
07/10/2024 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.4 F 0.0
2:00 170.5 F 2.6
4:00 176.1 F 6.5
6:00 196.8 F 10.9
8:00 215.8 F 6.1
10:00 225.8 F 10.1
12:00 230.9 F 8.3
14:00 240.2 F 12.5
16:00 240.8 F 11.3
18:00 253.2 F 17.5
20:00 263.9 F 24.0
22:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.7
1:00 272.9 F 30.1
2:00 271.4 F 29.7
3:00 272.3 F 30.1
4:00 271.5 F 29.6

Min 271.3 F 29.5
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 213.5 F 0.0
10:00 217.2 F 0.0
15:00 219.6 F 0.0
20:00 222.9 F 0.0
25:00 222.2 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

07/10/2024 01:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____