
Midmark M11 - v1.0.5
Total Cycles: 1916

Name: RITTER2
Sterilizer ID: V2313228
01/11/2023 08:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.5 F 0.0
2:00 79.8 F 0.0
4:00 89.0 F 0.0
6:00 117.4 F 0.0
8:00 162.4 F 0.1
10:00 190.3 F 0.3
12:00 202.6 F 3.4
14:00 215.8 F 6.8
16:00 226.7 F 10.6
18:00 231.1 F 7.8
20:00 239.8 F 11.6
22:00 238.0 F 9.8
24:00 237.8 F 9.0
26:00 250.4 F 15.5
28:00 262.2 F 22.3
30:00 272.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.7 F 29.2
2:00 271.6 F 28.5
3:00 272.8 F 29.3
4:00 271.8 F 28.5

Min 271.4 F 28.4
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.7 F 0.0
5:00 208.8 F 0.0
10:00 226.8 F 0.0
15:00 231.8 F 0.0
20:00 229.8 F 0.0
25:00 224.5 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:22
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:08:12
CYCLE COMPLETE

01/11/2023 09:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____