
Midmark M11 - v1.0.0
Total Cycles: 5099

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2024 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.9 F 0.0
2:00 120.3 F 0.0
4:00 160.2 F 0.3
6:00 179.5 F 0.6
8:00 203.9 F 1.2
10:00 216.1 F 5.2
12:00 226.1 F 8.7
14:00 236.5 F 13.1
16:00 236.2 F 9.5
18:00 239.4 F 9.5
20:00 245.9 F 13.2
22:00 244.0 F 12.0
24:00 256.9 F 18.8
26:00 268.1 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.5 F 28.5
2:00 272.1 F 28.6
3:00 272.1 F 28.3
4:00 272.5 F 28.9

Min 271.6 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.1 F 0.0
5:00 201.6 F 0.0
10:00 207.5 F 0.0
15:00 208.2 F 0.0
20:00 213.0 F 0.0
25:00 215.1 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 27:14
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:05:29
CYCLE COMPLETE

11/15/2024 04:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____