
Midmark M11 - v1.0.0
Total Cycles: 4742

Name: RITTER 1
Sterilizer ID: V2178198
07/17/2024 01:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.0 F 0.0
2:00 143.8 F 2.2
4:00 149.4 F 5.3
6:00 172.7 F 8.7
8:00 206.1 F 13.0
10:00 217.0 F 5.5
12:00 225.5 F 9.7
14:00 235.1 F 9.8
16:00 237.1 F 10.6
18:00 237.3 F 8.1
20:00 247.1 F 14.2
22:00 258.2 F 20.1
24:00 268.2 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.5
2:00 271.9 F 28.9
3:00 271.2 F 28.8
4:00 272.3 F 29.5

Min 271.1 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 207.1 F 0.0
10:00 206.8 F 0.0
15:00 216.6 F 0.0
20:00 238.2 F 0.0
25:00 236.0 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:04:06
CYCLE COMPLETE

07/17/2024 02:20 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____