
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002531
10:00 21/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:03
STERILIZING 3:04
136.4 C 223kPa 3:04
136.0 C 224kPa 3:34
136.0 C 222kPa 4:04
136.4 C 223kPa 4:34
136.4 C 226kPa 5:04
136.1 C 221kPa 5:34
136.4 C 225kPa 6:04
136.3 C 224kPa 6:34
136.3 C 224kPa 6:34
VENTING 6:34
AIR DRYING 7:26
CYCLE COMPLETE 37:26

Digital Signature #
2A17412F8645069A
